

Technical Specifications

Item No.		Statement of Compliance
1)	<p>1 UNIT DISSOLUTION APPARATUS</p> <p>Specifications:</p> <ul style="list-style-type: none"> - 6 stirrers, 230V according USP1, USP2, Manual, Filtering not included, Manual sampling by pipette, Outlet cannulas, including accessories - Fulfills all Pharmacopeia requirements for USP 1,2,5,6 dissolution methods - Innovative circular bath design for perfect visibility - Large touch screen, 100 methods storage, users rights, USB and LAN interface - Preferably made in Europe/USA - Delivery date 60 calendar days - With calibration certificate - Free Preventive Maintenance for 2 years - With training 	<p>Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a)(ii) and/or GCC Clause 2.1(a)(ii).</p>